

**PENNSYLVANIA DEPARTMENT OF HEALTH
BUREAU OF EMERGENCY MEDICAL SERVICES**

Vehicle #

BLS Ambulance Inspection Checklist

I. GENERAL INFORMATION:

Date Stickers: Yes No
Decals: Yes No

Name of EMS Agency:

Address:

(Primary Headquarters)

City

State

Zip

License Plate # :

Year:

Make:

Model:

Vehicle Identification # (VIN):

Date Inspected:

Affiliate # :

Regional EMS Council: -BLANK-

Mileage:

	YES	NO	N/A
Was a deficiency notification issued for this vehicle?			
Is a copy of the deficiency notification attached to this form?			
Is a reinspection required?			
VEHICLE/EQUIPMENT	PRESENT AND OPERATING	DEFICIENT	CORRECTED
Identified as Meeting the Fed KKK 1822 Specs			
Exterior Markings			
Audible Warning Signal			
Lights:			
Exterior			
Interior			
Dual Battery System			
Fire Extinguisher (2)(5# unit ABC dry chem or CO2)(Current Insp.)			
Power Supply			
Current Vehicle Inspection			
Current Vehicle Insurance			
Current Vehicle Registration			
Interior Requirements:			
Floor			
General Safety Concerns			
Patient Area Partition			
Storage Cabinets			
Bulky Items Secured			
Patient Litter Compliant With Manufacture Requirements			
Doors (side and rear gasket, latches and hinges)			
No Smoking /Oxygen Equipped Sign (2) (1 in front, 1 in rear)			
Fasten Seat Belts Sign (2) (1 in front, 1 in rear)			
Radio Equipment (meets regional comm. requirements)			
Installed Oxygen with min. 500L			
AMD Standard 003 for crashworthiness (min of 3 straps)			
with mounted O2 flow meter 0-25 lpm (1)			
Installed Suction (300mm/Hg in 4 sec.)			
Operational Heating/Cooling/Ventilation Equipment			
MEDICAL SUPPLIES/EQUIPMENT	PRESENT AND OPERATING	DEFICIENT	CORRECTED
Current Version of Statewide EMS Protocols			
Portable Suction Unit (1)(300mm/Hg in 4 sec.)			
Suction Catheters: (Sterile)			
Rigid (2)			
French (6 total) (1 each 6 & 8, 2-10 or 12, 2-14 or 16)			
Airways:			

Oropharyngeal (6 different sizes - to include one 0-1, one 2-3, & one 4-5)			
MEDICAL SUPPLIES/EQUIPMENT (CONT)	PRESENT AND OPERATING	DEFICIENT	CORRECTED
Nasopharyngeal (5 different sizes - to include one 16-24 Fr. & one 26-34 Fr.)			
Portable O2 cylinder with flow meter 0-25 lpm (1)			
With 300L & non-sparking wrench/tank opening device			
Secured in vehicle at all times			
Spare O2 cylinder (1) - secured in vehicle at all times			
Oxygen Delivery Devices:			
Nasal Cannulas (Adult & Pediatric- 1 each)			
High Concentration Masks (Adult, Infant, Pedi - 1 each)			
Pocket Mask with One-Way Valve & O2 port (1)			
Humidifier bottle (1)			
Bag Valve Mask Devices-(1)Adult & (1)Pedi (450-700cc)			
Masks to include Adult, neonatal, infant & child			
Sphygmomanometer			
(Child, Adult & Thigh(Lg)-1 each or interchangeable cuffs)			
Stethoscope (Adult & Pediatric - 1 each)			
Penlight (1)			
Dressings:			
Multi-Trauma (10" x 30") (4)			
Occlusive (3" x 4") (4)			
Sterile Gauze Pads (3" x 3") (25)			
Soft Self Adhering (6 rolls)			
Adhesive Tape (4 rolls assort., 1 must be hypoallergenic)			
Bandage Shears (1)			
Commercial "Tactical" Tourniquet (1)			
Immobilization Devices:			
Lateral Cervical Spine Device (1)			
Long Spine Board (1)			
Short Spine Board (1)			
Rigid/Semi Rigid Neck Immobilizers			
(S, M, L, & Pedi.-1 each or Multi -size (3 & 1 Pedi)			
Straps 9' (5)(May sub spider straps or speed clips for 3)			
Folding Litter/Collapsible Device (1)			
Splinting Devices:			
Traction Splint (Adult & Child - 1 each or Comb)			
Upper & Lower Extremity Splints (2 each)			
Pediatric Equipment Sizing Tape/Chart (current)- BLS			
Sterile Water/Normal Saline- 2 liters			
Cold Packs, Chemical (4)			
Heat Packs, Chemical (4)			
Triangular Bandages (8)			
Sterile OB Kit (2)			
Separate Bulb Syringe (1) Sterile			
Thermal Blanket-Silver Swaddler or roll of Sterile Foil (1)			
Sterile Burn Sheets (4' x 4') (2)			
Pillow (1)			
Blankets (2)			
Sheets (4)			
Pillow Cases (2)			
Towels (4)			
Disposable Tissues (1 box)			
Emesis Container (1)			
Bedpan (1)			
Urinal (1)			
Disposable Paper Drinking Cups (3 oz.) (4)			
Emergency BLS Jump Kit (1)			

	PRESENT AND OPERATING	DEFICIENT	CORRECTED
MEDICAL SUPPLIES/EQUIPMENT (CONT)			
Thermometer (1) elec, dig, non-tympanic			
Instant Glucose (45 grams-40% dextrose-d-glucose gel)			
Lubrication (2cc or Larger tube) sterile water soluble (2)			
Epinephrine Auto-injector, Adult & Pedi (2 each)(opt. BLS)			
CPAP Ventilation - portable equipment (opt. BLS)			
Electronic Glucose Meter (1) (opt. BLS)			
Naloxone (opt. BLS)			
Pulse Oximetry			
Aspirin 81 mg			
AED (with 2 set of pads) at least (1) set of pads need to be adult			

	PRESENT AND OPERATING	DEFICIENT	CORRECTED
PERSONAL PROTECTIVE EQUIPMENT			
Hand light (2)			
Hazard Warning Device (3)			
High-visibility safety apparel (1 per crew member)			
Helmet (1 per crew member)			
Gloves (leather) (1 pair per crew member)			
Eye Protection - Goggles (1 pair per crew member)			
Regional Approved Triage Tags (20)			
DOT Emergency Response Guide (1) - Current Ed.			

	PRESENT	DEFICIENT	CORRECTED
PERSONAL PROTECTION EQUIPMENT			
Eye Protection			
Face Mask*			
Gown/Coat*			
Surgical Cap/Foot Coverings*			
Exam Gloves*			
Biological Waste Container			
Red Bags for Waste Container			
Sharps container - secured			
Face Mask (N-95 one per crew member)*			
Hand Disinfectant/cleaner - Non-water (1 container)			
* Disposable -one set/pair per responding crewmember			

Electronic Deficiency Form Completed Yes N/A

Digital Images Captured Yes N/A

Vehicle Placed Out of Service (Per I.B. 2013-001) Yes N/A

Inspected By: _____ (Printed Name)

Signature: _____

Date Forwarded to BEMS: _____